


# Polymer Systems Technology Limited

UK & Ireland Distributor



© 2008 - Polymer Systems Technology Limited <sup>TM</sup>  
Unit 2. Network 4. Cressex Business Park,  
Lincoln Road, High Wycombe, Bucks. HP12 3RF  
Phone +44 (0) 1494 446610  
Fax: +44 (0) 1494 528611  
Web: <http://www.siliconepolymers.co.uk>  
Email: [sales@silicone-polymers.co.uk](mailto:sales@silicone-polymers.co.uk)



# MED-6019

Optically Clear, High Durometer Silicone Elastomer

## Product Profile

### Description

- Flowable, two-part, optically clear, high durometer silicone
- Cures at low temperatures
- Optional post-cure increases durometer
- 1:1 Mix Ratio (Part A:B)

### Applications

- Use for potting and encapsulating electronic devices and for making molded parts
- Useful in any application requiring a flowable, high durometer material

**NuSil Technology's MED-6019 is a restricted product. It shall not be considered for use in human implantation for a period of greater than 29 days.**

Typical Properties	Result	Metric Conv.	ASTM	NT-TM
<b>Uncured:</b>				
Appearance	Clear	-	D2090	002
Viscosity, Part A	33,000 cP	33,000 mPas	D1084, D2196	001
Viscosity, Part B	7,500 cP	7,500 mPas	D1084, D2196	001
Work Time	1.5 hours	-	-	008
<b>Cured : 1 hour @ 40°C (104°F)</b>				
Durometer, Type A	75	-	D2240	006
Refractive Index	1.41	-	D1747, D1218	018
Tensile Strength	1,100 psi	7.6 MPa	D412, D882	007
Elongation	100%	-	D412, D882	007
Tear Strength	50 ppi	8.8 kN/m	D624	009
<b>Post-Cured: 1 hour @ 150°C (302°F)</b>				
Specific Gravity	1.07	-	D792	003
Durometer, Type A	85	-	D2240	006
Tensile Strength	500 psi	3.4 MPa	D412, D882	007
Elongation	25%	-	D412, D882	007
Cytotoxicity	Pass	-	-	061

### Instructions for Use

#### Mixing

Thoroughly mix Parts A and B in a 1:1 ratio by weight.

#### Vacuum Deaeration

Remove air entrapped during mixing by common vacuum de-aeration procedure, observing all applicable safety precautions. Slowly apply full vacuum to a container rated for use and at least four times the volume of material being de-aired. Hold vacuum until bulk deaeration is complete.

#### Substrate Consideration

MED-6019 will cure in contact with most materials common to biomedical assemblies. Exceptions include sulfur-cured organic rubbers, latex, chlorinated rubbers, some RTV silicones and unreacted residues of some curing agents. Units being encapsulated or potted should be clean and free of surface contaminants. Containers and dispensers being used with MED-6019 should also be clean and dry. Washing all containers with clean solvent or volatilizing the contaminants by heating can usually prevent cure inhibition.

### Packaging

50 ml SxS Kit  
 400 ml SxS Kit  
 2 Pint Kit (910 g)  
 2 Gallon Kit (7.28 kg)  
 10 Gallon Kit (36.4 kg)

### Warranty

12 Months

## **FDA Master Access File**

A Master Access File for MED-6019 will be filed with the U.S. Food and Drug Administration. Customers interested in authorization to reference the Master Access File must contact NuSil Technology.

## **Warnings About Product Safety**

NuSil Technology believes the information and the data contained herein are accurate and reliable. However, the user is responsible to determine the material's suitability and safety of use. NuSil Technology cannot know each application's specific requirements and hereby notifies the user that it has not tested or determined this material's suitability or safety for use in any application. The user is responsible to adequately test and determine the safety and suitability for their application and NuSil Technology makes no warranty concerning fitness for any use or purpose. NuSil Technology has completed no testing to establish safety of use in any medical application.

NuSil Technology has tested this material only to determine if the product meets the applicable specifications. (Please contact NuSil Technology for assistance and recommendations when establishing specifications.) When considering the use of NuSil Technology products in a particular application, review the latest Material Safety Data Sheet and contact NuSil Technology with any questions about product safety information.

Do not use any chemical in a food, drug, cosmetic, or medical application or process until having determined the safety and legality of the use. The user is responsible to meet the requirements of the U.S. Food and Drug Administration (FDA) and any other regulatory agencies. Before handling any other materials mentioned in the text, obtain available product safety information and take the necessary steps to ensure safety of use.

## **Specifications**

Do not use the typical properties shown in this technical profile as a basis for preparing specifications. Please contact NuSil Technology for assistance and recommendations in establishing particular specifications.

## **Patent Warning**

NuSil Technology disclaims any expressed or implied warranty against the infringement of any patent. NuSil Technology does not warrant the use or sale of the products described herein will not infringe the claims of any United States' or other country's patents covering the product itself, its use in combination with other products or its use in the operation of any process.

## **Warranty Information**

NuSil Technology's warranty period is 12 months from the date of shipment when stored below 40°C in original unopened containers. Unless NuSil Technology provides a specific written warranty of fitness for a particular use, NuSil Technology's sole warranty is that the product will meet NuSil Technology's then current specification. NuSil Technology specifically disclaims any other expressed or implied warranty, including warranties of merchantability and fitness for use. The exclusive remedy and NuSil Technology's sole liability for breach of warranty is limited to refund of purchase price or replacement of any product shown to be other than as warranted. NuSil Technology expressly disclaims any liability for incidental or consequential damages.